



FACTORY ACT THERMOMETER



Operating Instructions

INSTRUMENT OPERATION - For the most accurate recording of ambient temperatures, the thermometer should be positioned away from direct sunlight and draughts.

THERMOMETER TEMPERATURES - The UK Workplace Health, Safety and Welfare Regulations 1992 requires the temperature in the workroom to be at least 16 °C/61 °F unless much of the work involves severe physical effort in which case the temperature should be at least 13 °C/55 °F.

FR - FONCTIONNEMENT DE L'INSTRUMENT - Pour un enregistrement le plus précis possible des températures ambiantes, le thermomètre doit être positionné loin d'une source de lumière directe ou de courants d'air.

DE - GEBRAUCHSANWEISUNGEN - Zur genauestmöglichen Aufzeichnung der Umgebungstemperaturen sollte das Thermometer weder direkter Sonneneinstrahlung noch Zugluft ausgesetzt sein.

IT - FUNZIONAMENTO DELLO STRUMENTO - Per una registrazione ottimale della temperatura ambientale, posizionare il termometro lontano dalla luce diretta del sole e mantenerlo protetto dalle correnti d'aria.

PT - MANUSEAMENTO DO INSTRUMENTO - Para registar a temperatura ambiente com maior precisão, o termómetro não deve ficar exposto à luz directa do sol, nem a correntes de ar.

ES - FUNCIONAMIENTO DEL INSTRUMENTO - Para obtener registros exactos de la temperatura ambiente, coloque el termómetro lejos de la luz solar directa y en un entorno seco.

GUARANTEE - This instrument carries a one-year guarantee against defects in either components or workmanship. During this period, products that prove to be defective will, at the discretion of ETI, be either repaired or replaced without charge. The product guarantee does not cover damage caused by fair wear and tear, abnormal storage conditions, incorrect use, accidental misuse, abuse, neglect, misapplication or modification. Full details of liability are available within ETI's Terms & Conditions of Sale at etiltld.com. In line with our policy of continuous development, we reserve the right to amend our product specification without prior notice.

Supplied by
Electronic Temperature Instruments Ltd
Worthing · West Sussex · BN14 8HQ
01903 202151 · sales@etiltld.com · etiltld.com

